

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076163.D
 Acq On : 09 Jan 2023 18:04
 Operator : JC\MD
 Sample : 01042-05 500X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HYATT-SILICA

Quant Time: Jan 10 01:10:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	325736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	518753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	442008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	181837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	171045	48.267	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.540%
35) Dibromofluoromethane	8.177	113	156562	48.734	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	10.571	98	593622	49.421	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.853	95	184677	46.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.840%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	7663	6.273	ug/l #	84
40) Benzene	8.612	78	391479	29.732	ug/l	98
52) Toluene	10.635	92	302546	35.281	ug/l	99
67) Ethyl Benzene	11.965	91	27447	1.793	ug/l	93
68) m/p-Xylenes	12.070	106	55532	9.163	ug/l	100
69) o-Xylene	12.400	106	19369	3.230	ug/l	100
95) Naphthalene	15.641	128	9614	3.107	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076163.D
Acq On : 09 Jan 2023 18:04
Operator : JC\MD
Sample : 01042-05 500X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HYATT-SILICA

Quant Time: Jan 10 01:10:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

