

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076165.D
 Acq On : 09 Jan 2023 18:52
 Operator : JC\MD
 Sample : 01053-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-201

Quant Time: Jan 10 01:11:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	337617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	540663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	481042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	181438	49.398	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.800%
35) Dibromofluoromethane	8.177	113	161990	48.381	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	10.571	98	635042	50.727	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.853	95	197452	47.621	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.240%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	34396	27.166	ug/l	98
20) Methylene Chloride	5.300	84	8738	1.965	ug/l #	95
37) Ethyl Acetate	7.571	43	32306	9.043	ug/l	98
43) Isopropyl Acetate	8.694	43	246602	41.844	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076165.D
Acq On : 09 Jan 2023 18:52
Operator : JC\MD
Sample : 01053-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-201

Quant Time: Jan 10 01:11:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

