

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
 Data File : VN076169.D
 Acq On : 09 Jan 2023 20:28
 Operator : JC\MD
 Sample : 01086-06 200X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1844

Quant Time: Jan 10 01:13:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	315975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	501495	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	451896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	173326	50.422	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.840%		
35) Dibromofluoromethane	8.177	113	153993	49.584	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.160%		
50) Toluene-d8	10.571	98	584543	50.340	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.680%		
62) 4-Bromofluorobenzene	12.853	95	185866	48.328	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.660%		
Target Compounds						
						Qvalue
52) Toluene	10.630	92	2120	0.256	ug/l	78
68) m/p-Xylenes	12.065	106	1309	0.211	ug/l	80
95) Naphthalene	15.641	128	3532	2.501	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010923\
Data File : VN076169.D
Acq On : 09 Jan 2023 20:28
Operator : JC\MD
Sample : 01086-06 200X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1844

Quant Time: Jan 10 01:13:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

