

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
 Data File : VN085403.D
 Acq On : 09 Jan 2025 12:43
 Operator : JC\MD
 Sample : VN0109WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0109WBL01

Quant Time: Jan 10 00:16:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	182405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	342003	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	292931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119086	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150556	57.870	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.740%
35) Dibromofluoromethane	8.165	113	115178	53.983	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.960%
50) Toluene-d8	10.559	98	408966	52.235	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.460%
62) 4-Bromofluorobenzene	12.847	95	123089	47.338	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
Data File : VN085403.D
Acq On : 09 Jan 2025 12:43
Operator : JC\MD
Sample : VN0109WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0109WBL01

Quant Time: Jan 10 00:16:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

