

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
 Data File : VN085410.D
 Acq On : 09 Jan 2025 15:39
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 10 00:22:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	161438	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302815	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	265400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105816	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	135332	58.774	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	117.540%	
35) Dibromofluoromethane	8.165	113	103508	54.792	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	109.580%	
50) Toluene-d8	10.565	98	371965	53.657	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	107.320%	
62) 4-Bromofluorobenzene	12.847	95	114258	49.629	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	99.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
Data File : VN085410.D
Acq On : 09 Jan 2025 15:39
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 10 00:22:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

