

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
 Data File : VN085412.D
 Acq On : 09 Jan 2025 16:27
 Operator : JC\MD
 Sample : Q1025-01
 Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GREENBROOK-COMP

Quant Time: Jan 10 00:24:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	141304	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	259943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	215785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	89132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	110796	54.975	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.940%			
35) Dibromofluoromethane	8.159	113	84819	52.304	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.600%			
50) Toluene-d8	10.559	98	302173	50.779	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.560%			
62) 4-Bromofluorobenzene	12.847	95	89674	45.374	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.740%			
Target Compounds						
8) Diethyl Ether	3.953	74	888	0.898	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
Data File : VN085412.D
Acq On : 09 Jan 2025 16:27
Operator : JC\MD
Sample : Q1025-01
Misc : 4.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GREENBROOK-COMP

Quant Time: Jan 10 00:24:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

