

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
 Data File : VN085414.D
 Acq On : 09 Jan 2025 17:35
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 10 00:25:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 28 01:36:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	185275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	354078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144515	54.688	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	109.380%	
35) Dibromofluoromethane	8.165	113	113378	51.327	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.660%	
50) Toluene-d8	10.565	98	421209	51.964	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.920%	
62) 4-Bromofluorobenzene	12.847	95	143689	53.376	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	106.760%	
Target Compounds						
95) Naphthalene	15.641	128	8295	1.228	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN010925\
Data File : VN085414.D
Acq On : 09 Jan 2025 17:35
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 10 00:25:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122724W.M
Quant Title : SW846 8260
QLast Update : Sat Dec 28 01:36:16 2024
Response via : Initial Calibration

