

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011019\
 Data File : VN053338.D
 Acq On : 10 Jan 2019 13:41
 Operator : MD\SY
 Sample : K1103-01 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-WW

Quant Time: Jan 11 06:17:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 28 10:33:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	154353	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	875416	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	776753	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	294832	30.07	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.23%
60) 4-Bromofluorobenzene	12.40	95	341743	28.75	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	95.83%
63) Toluene-d8	10.09	98	1034172	29.55	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.50%

Target Compounds

					Ovalue
14) Acrylonitrile	4.99	53	19501	6.218	ug/l # 88
15) Acetone	3.82	58	75929	113.924	ug/l 91
25) Chloroform	7.37	83	986268	58.160	ug/l 100
41) Bromodichloromethane	9.40	83	10444	0.826	ug/l 96
59) 2-Hexanone	10.75	43	5218	1.021	ug/l 99
91) Naphthalene	15.14	128	11407	0.611	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011019\
Data File : VN053338.D
Acq On : 10 Jan 2019 13:41
Operator : MD\SY
Sample : K1103-01 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EFF-WW

Quant Time: Jan 11 06:17:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 28 10:33:50 2018
Response via : Initial Calibration

