

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011019\
 Data File : VN053341.D
 Acq On : 10 Jan 2019 15:01
 Operator : MD\SY
 Sample : K1005-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-010918-B

Quant Time: Jan 11 06:24:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1028553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1596871	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1509702	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	632875	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	510651	47.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.44%	
35) Dibromofluoromethane	7.59	113	452928	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
50) Toluene-d8	10.09	98	1980994	50.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.02%	
62) 4-Bromofluorobenzene	12.40	95	693480	50.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.02%	
Target Compounds						
16) Acetone	3.82	43	25555	7.757	ug/l	97
20) Methylene Chloride	4.55	84	569671	49.567	ug/l	98
43) Isopropyl Acetate	8.17	43	41978	2.198	ug/l #	80
80) 1,3,5-Trimethylbenzene	12.73	105	85144	2.243	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	118828	3.073	ug/l	99
95) Naphthalene	15.13	128	109299	9.823	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011019\
Data File : VN053341.D
Acq On : 10 Jan 2019 15:01
Operator : MD\SY
Sample : K1005-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
HR-06-010918-B

Quant Time: Jan 11 06:24:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

