

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011019\
 Data File : VN053349.D
 Acq On : 10 Jan 2019 19:02
 Operator : MD\SY
 Sample : K1120-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: Jan 11 06:41:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 28 10:33:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	140680	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	805047	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	684557	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	281336	31.48	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.93%
60) 4-Bromofluorobenzene	12.40	95	289917	27.68	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.27%
63) Toluene-d8	10.09	98	934533	30.30	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011019\
Data File : VN053349.D
Acq On : 10 Jan 2019 19:02
Operator : MD\SY
Sample : K1120-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: Jan 11 06:41:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 28 10:33:50 2018
Response via : Initial Calibration

