

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
 Data File : VN070539.D
 Acq On : 10 Jan 2022 12:44
 Operator : JC/MD
 Sample : VN0110WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0110WBL02

Quant Time: Jan 10 17:27:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	1112447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1820305	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1612811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	558079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	773246	48.343	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.680%
35) Dibromofluoromethane	8.021	113	551351	50.060	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.120%
50) Toluene-d8	10.440	98	2185968	48.585	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.180%
62) 4-Bromofluorobenzene	12.731	95	815249	50.168	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
Data File : VN070539.D
Acq On : 10 Jan 2022 12:44
Operator : JC/MD
Sample : VN0110WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0110WBL02

Quant Time: Jan 10 17:27:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

