

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
 Data File : VN070544.D
 Acq On : 10 Jan 2022 15:38
 Operator : JC/MD
 Sample : N1012-07DL 40X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31239DL

Quant Time: Jan 11 05:35:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	981940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1653730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1458200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	496173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	521183	36.915	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	73.840%
35) Dibromofluoromethane	8.021	113	371504	37.129	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	74.260%
50) Toluene-d8	10.440	98	1451259	35.505	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	71.000%
62) 4-Bromofluorobenzene	12.731	95	513233	34.764	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	69.520%
Target Compounds						
16) Acetone	4.288	43	413225	45.087	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
Data File : VN070544.D
Acq On : 10 Jan 2022 15:38
Operator : JC/MD
Sample : N1012-07DL 40X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31239DL

Quant Time: Jan 11 05:35:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

