

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
 Data File : VN070552.D
 Acq On : 10 Jan 2022 19:00
 Operator : JC/MD
 Sample : N1015-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30915

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/11/2022
 Supervised By : Mahesh Dadoda 01/12/2022

Quant Time: Jan 11 06:06:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1014733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1716087	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1546932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	596258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	717200	49.157	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.320%		
35) Dibromofluoromethane	8.021	113	503797	48.520	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.040%		
50) Toluene-d8	10.440	98	2101981	49.556	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.120%		
62) 4-Bromofluorobenzene	12.728	95	816717	53.311	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	106.620%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.369	59	10915m	4.493	ug/l	
13) Acrolein	4.049	56	12974	6.582	ug/l	85
16) Acetone	4.285	43	2498541	263.806	ug/l	98
18) Methyl Acetate	4.857	43	67809	2.475	ug/l	91
25) 2-Butanone	7.335	43	247480	20.752	ug/l	90
40) Benzene	8.458	78	121253	2.303	ug/l	95
52) Toluene	10.505	92	40752	1.279	ug/l	97
68) m/p-Xylenes	11.945	106	12520	0.572	ug/l	95
69) o-Xylene	12.280	106	5580	0.256	ug/l	87
95) Naphthalene	15.504	128	38275	3.155	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
Data File : VN070552.D
Acq On : 10 Jan 2022 19:00
Operator : JC/MD
Sample : N1015-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30915

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/11/2022
Supervised By : Mahesh Dadoda 01/12/2022

Quant Time: Jan 11 06:06:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

