

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
 Data File : VN070553.D
 Acq On : 10 Jan 2022 19:26
 Operator : JC/MD
 Sample : N1015-07DL 200X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20986DL

Quant Time: Jan 11 06:06:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	967318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1656975	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1471150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	546403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	746536	53.676	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.360%	
35) Dibromofluoromethane	8.021	113	520990	51.966	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.940%	
50) Toluene-d8	10.440	98	2167162	52.915	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.840%	
62) 4-Bromofluorobenzene	12.731	95	803794	54.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.680%	
Target Compounds						
						Qvalue
13) Acrolein	4.041	56	12118	6.449	ug/l	# 14
16) Acetone	4.283	43	3373872	373.688	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011022\
Data File : VN070553.D
Acq On : 10 Jan 2022 19:26
Operator : JC/MD
Sample : N1015-07DL 200X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20986DL

Quant Time: Jan 11 06:06:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

