

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011119\
 Data File : VN053358.D
 Acq On : 11 Jan 2019 11:35
 Operator : MD\SY
 Sample : PB116316ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116316ZHE#01

Quant Time: Jan 12 01:09:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	976337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1511192	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1433407	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	616836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	495783	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
35) Dibromofluoromethane	7.59	113	431354	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.09	98	1898238	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.40	95	661581	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
Target Compounds						
16) Acetone	3.82	43	19378	5.743	ug/l	95
20) Methylene Chloride	4.56	84	15123	1.386	ug/l	91
43) Isopropyl Acetate	8.17	43	42385	2.345	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011119\
Data File : VN053358.D
Acq On : 11 Jan 2019 11:35
Operator : MD\SY
Sample : PB116316ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116316ZHE#01

Quant Time: Jan 12 01:09:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

