

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011119\
 Data File : VN053362.D
 Acq On : 11 Jan 2019 13:22
 Operator : JC/SP
 Sample : PB116316ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116316ZHE#05

Quant Time: Jan 12 01:23:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1002400	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1558044	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1496423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	666171	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	510107	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
35) Dibromofluoromethane	7.59	113	444186	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
50) Toluene-d8	10.09	98	1959986	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
62) 4-Bromofluorobenzene	12.40	95	690131	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
Target Compounds						
16) Acetone	3.82	43	21086	6.222	ug/l	97
20) Methylene Chloride	4.56	84	16067	1.434	ug/l	93
43) Isopropyl Acetate	8.17	43	46797	2.511	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011119\
Data File : VN053362.D
Acq On : 11 Jan 2019 13:22
Operator : JC/SP
Sample : PB116316ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116316ZHE#05

Quant Time: Jan 12 01:23:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

