

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011119\
 Data File : VN053365.D
 Acq On : 11 Jan 2019 14:43
 Operator : JC/SP
 Sample : PB116316ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB116316ZHE#08

Quant Time: Jan 12 01:32:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1048875	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1658732	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1559484	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	647240	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	539277	49.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.84%	
35) Dibromofluoromethane	7.59	113	471400	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
50) Toluene-d8	10.09	98	2058792	50.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.08%	
62) 4-Bromofluorobenzene	12.40	95	710410	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
Target Compounds						
16) Acetone	3.82	43	15619	3.746	ug/l	98
20) Methylene Chloride	4.55	84	11874	1.013	ug/l	97
43) Isopropyl Acetate	8.17	43	51851	2.613	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011119\
Data File : VN053365.D
Acq On : 11 Jan 2019 14:43
Operator : JC/SP
Sample : PB116316ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB116316ZHE#08

Quant Time: Jan 12 01:32:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

