

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011119\
 Data File : VN053372.D
 Acq On : 11 Jan 2019 18:00
 Operator : JC/SP
 Sample : K1119-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 DS-A

Quant Time: Jan 12 02:33:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1010269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1575399	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1496290	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	688078	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	514331	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	7.59	113	451076	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
50) Toluene-d8	10.09	98	1970147	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	12.40	95	696329	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
Target Compounds						
					Qvalue	
16) Acetone	3.82	43	40308	13.823	ug/l	94
20) Methylene Chloride	4.55	84	224388	19.877	ug/l	97
43) Isopropyl Acetate	8.17	43	45438	2.411	ug/l #	78
68) m/p-Xylenes	11.62	106	37234	1.793	ug/l	98
69) o-Xylene	11.95	106	27694	1.362	ug/l	98
80) 1,3,5-Trimethylbenzene	12.73	105	88328	2.140	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	370545	8.814	ug/l	100
95) Naphthalene	15.13	128	289369	23.919	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011119\
Data File : VN053372.D
Acq On : 11 Jan 2019 18:00
Operator : JC/SP
Sample : K1119-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
DS-A

Quant Time: Jan 12 02:33:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

