

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011119\
 Data File : VN053375.D
 Acq On : 11 Jan 2019 19:20
 Operator : JC/SP
 Sample : K1119-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 DS-D

Quant Time: Jan 12 02:44:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 07 09:34:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1037586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1605432	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1520786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	691626	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	519473	48.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.24%	
35) Dibromofluoromethane	7.59	113	459230	50.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.50%	
50) Toluene-d8	10.09	98	2007525	50.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.84%	
62) 4-Bromofluorobenzene	12.40	95	715331	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
Target Compounds						
16) Acetone	3.82	43	26751	8.134	ug/l	95
20) Methylene Chloride	4.55	84	38330	3.306	ug/l	94
43) Isopropyl Acetate	8.17	43	47611	2.479	ug/l #	78
68) m/p-Xylenes	11.62	106	78802	3.733	ug/l	97
69) o-Xylene	11.95	106	37243	1.802	ug/l	100
80) 1,3,5-Trimethylbenzene	12.73	105	43547	1.050	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	182515	4.319	ug/l	100
95) Naphthalene	15.14	128	31979	2.630	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011119\
Data File : VN053375.D
Acq On : 11 Jan 2019 19:20
Operator : JC/SP
Sample : K1119-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
DS-D

Quant Time: Jan 12 02:44:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N010519W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 07 09:34:27 2019
Response via : Initial Calibration

