

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065460.D
 Acq On : 11 Jan 2021 11:43
 Operator : JC/MD
 Sample : VN0111WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0111WBL01

Quant Time: Jan 12 00:05:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	228153	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	367311	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	390790	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	154366	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	123808	48.07	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.14%
35) Dibromofluoromethane	7.553	113	114405	51.63	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.26%
50) Toluene-d8	10.061	98	463766	55.84	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.68%
62) 4-Bromofluorobenzene	12.376	95	175072	52.96	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
Data File : VN065460.D
Acq On : 11 Jan 2021 11:43
Operator : JC/MD
Sample : VN0111WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0111WBL01

Quant Time: Jan 12 00:05:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

