

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065463.D
 Acq On : 11 Jan 2021 13:12
 Operator : JC/MD
 Sample : M1022-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41038

Quant Time: Jan 12 00:07:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	218955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	361817	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	388193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	170205	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	117253	47.44	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 94.88%			
35) Dibromofluoromethane	7.553	113	109268	50.06	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.12%			
50) Toluene-d8	10.061	98	444703	54.36	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.72%			
62) 4-Bromofluorobenzene	12.376	95	178907	54.75	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 109.50%			
Target Compounds						Qvalue
11) Tert butyl alcohol	4.778	59	7370	18.22	ug/l #	91
16) Acetone	3.791	43	144893	134.73	ug/l	94
17) Carbon Disulfide	4.010	76	9914	1.80	ug/l	98
18) Methyl Acetate	4.296	43	160512	72.32	ug/l	99
25) 2-Butanone	6.807	43	19852	14.31	ug/l	100
52) Toluene	10.126	92	6503	1.05	ug/l	95
68) m/p-Xylenes	11.592	106	2355	0.44	ug/l	94
69) o-Xylene	11.920	106	1188	0.23	ug/l	99
80) 1,3,5-Trimethylbenzene	12.707	105	2368	0.21	ug/l	96
84) 1,2,4-Trimethylbenzene	13.013	105	12764	1.19	ug/l	74
89) n-Butylbenzene	13.585	91	3177	0.32	ug/l #	73
95) Naphthalene	15.103	128	11171	4.55	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
Data File : VN065463.D
Acq On : 11 Jan 2021 13:12
Operator : JC/MD
Sample : M1022-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41038

Quant Time: Jan 12 00:07:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

