

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065465.D
 Acq On : 11 Jan 2021 14:01
 Operator : JC/MD
 Sample : M1058-02DL 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB-2021-WW-000-05DL

Quant Time: Jan 12 00:07:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	219962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	361486	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	383985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	156784	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	119978	48.32	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.64%
35) Dibromofluoromethane	7.553	113	111126	50.95	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.90%
50) Toluene-d8	10.061	98	446072	54.57	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.14%
62) 4-Bromofluorobenzene	12.376	95	173367	53.26	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.52%
Target Compounds						
					Qvalue	
30) Chloroform	7.338	83	1628	0.41	ug/l	91
44) Trichloroethene	8.804	130	4487	1.59	ug/l	93
64) Tetrachloroethene	10.601	164	130246	34.93	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
Data File : VN065465.D
Acq On : 11 Jan 2021 14:01
Operator : JC/MD
Sample : M1058-02DL 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB-2021-WW-000-05DL

Quant Time: Jan 12 00:07:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

