

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065466.D
 Acq On : 11 Jan 2021 14:26
 Operator : JC/MD
 Sample : M1039-01 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41030

Quant Time: Jan 12 00:07:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	211939	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	360066	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	377657	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	161402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	118656	49.59	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 99.18%			
35) Dibromofluoromethane	7.553	113	108921	50.14	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.28%			
50) Toluene-d8	10.061	98	435472	53.49	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.98%			
62) 4-Bromofluorobenzene	12.376	95	170946	52.77	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 105.54%			
Target Compounds					Qvalue	
16) Acetone	3.792	43	83667	75.84	ug/l	94
25) 2-Butanone	6.811	43	3964	2.95	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
Data File : VN065466.D
Acq On : 11 Jan 2021 14:26
Operator : JC/MD
Sample : M1039-01 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41030

Quant Time: Jan 12 00:07:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

