

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065474.D
 Acq On : 11 Jan 2021 17:44
 Operator : JC/MD
 Sample : M1038-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CT-SP

Quant Time: Jan 12 00:10:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	218668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	360846	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	393887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	163356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	119502	48.41	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery =	96.82%		
35) Dibromofluoromethane	7.553	113	109865	50.47	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery =	100.94%		
50) Toluene-d8	10.061	98	453248	55.55	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery =	111.10%		
62) 4-Bromofluorobenzene	12.376	95	180248	55.26	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery =	110.52%		
Target Compounds						
						Qvalue
20) Methylene Chloride	4.508	84	8445	3.51	ug/l	96
43) Isopropyl Acetate	8.135	43	17106	3.05	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065474.D
 Acq On : 11 Jan 2021 17:44
 Operator : JC/MD
 Sample : M1038-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CT-SP

Quant Time: Jan 12 00:10:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

