

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
 Data File : VN065482.D
 Acq On : 11 Jan 2021 21:02
 Operator : JC/MD
 Sample : M1058-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB-2021-WW-000-05

Quant Time: Jan 12 00:12:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 22 15:42:18 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	213796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	353881	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	380010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	165647	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.990	65	115330	47.78	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.56%
35) Dibromofluoromethane	7.553	113	106778	50.01	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.02%
50) Toluene-d8	10.061	98	435162	54.38	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.76%
62) 4-Bromofluorobenzene	12.376	95	173769	54.41	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.82%
Target Compounds						
					Qvalue	
16) Acetone	3.795	43	26078	15.66	ug/l #	87
25) 2-Butanone	6.810	43	10495	7.75	ug/l	95
27) cis-1,2-Dichloroethene	6.788	96	1575	0.66	ug/l	87
30) Chloroform	7.335	83	28779	7.40	ug/l	98
40) Benzene	8.003	78	12548	1.31	ug/l	99
44) Trichloroethene	8.804	130	82475	29.85	ug/l	93
47) Bromodichloromethane	9.373	83	6253	1.75	ug/l #	93
59) 2-Hexanone	10.727	43	4440	1.95	ug/l	90
64) Tetrachloroethene	10.605	164	2395219	649.04	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011121\
Data File : VN065482.D
Acq On : 11 Jan 2021 21:02
Operator : JC/MD
Sample : M1058-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB-2021-WW-000-05

Quant Time: Jan 12 00:12:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122220W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 22 15:42:18 2020
Response via : Initial Calibration

