

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011122\
 Data File : VN070567.D
 Acq On : 11 Jan 2022 13:54
 Operator : JC/MD
 Sample : N1050-01ME
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-236ME

Quant Time: Jan 12 06:00:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	898743	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1763283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2216274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	968092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	640333	49.553	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.100%
35) Dibromofluoromethane	8.016	113	485945	45.548	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.100%
50) Toluene-d8	10.438	98	2408273	55.257	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	110.520%
62) 4-Bromofluorobenzene	12.728	95	1123278	71.359	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	142.720%#
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	14635	1.745	ug/l	93
25) 2-Butanone	7.340	43	83290	7.886	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011122\
Data File : VN070567.D
Acq On : 11 Jan 2022 13:54
Operator : JC/MD
Sample : N1050-01ME
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-236ME

Quant Time: Jan 12 06:00:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

