

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011122\
 Data File : VN070569.D
 Acq On : 11 Jan 2022 15:01
 Operator : JC/MD
 Sample : N1050-01ME
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-236ME

Quant Time: Jan 12 06:01:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.078	168	997200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1668213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1483183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	606423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	563030	39.269	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	78.540%		
35) Dibromofluoromethane	8.013	113	390841	38.722	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	77.440%		
50) Toluene-d8	10.438	98	1673741	40.592	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	81.180%		
62) 4-Bromofluorobenzene	12.731	95	687520	46.165	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.340%		
Target Compounds						Qvalue
25) 2-Butanone	7.340	43	112790	9.624	ug/l	92
52) Toluene	10.502	92	45073	1.455	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011122\
Data File : VN070569.D
Acq On : 11 Jan 2022 15:01
Operator : JC/MD
Sample : N1050-01ME
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-236ME

Quant Time: Jan 12 06:01:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

