

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011122\
 Data File : VN070573.D
 Acq On : 11 Jan 2022 17:26
 Operator : JC/MD
 Sample : N1098-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14-G-WATER

Quant Time: Jan 12 06:02:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		8.083	168	1122955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		8.965	114	1847050	50.000	ug/l	0.00
63) Chlorobenzene-d5		11.741	117	1647080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		13.672	152	625973	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		8.434	65	820195	50.799	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.600%	
35) Dibromofluoromethane		8.019	113	586058	52.441	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.880%	
50) Toluene-d8		10.438	98	2399243	52.553	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.100%	
62) 4-Bromofluorobenzene		12.728	95	917144	55.621	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.240%	
Target Compounds							
16) Acetone		4.285	43	36741	3.505	ug/l	100
19) Methyl tert-butyl Ether		5.610	73	15997	0.377	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011122\
Data File : VN070573.D
Acq On : 11 Jan 2022 17:26
Operator : JC/MD
Sample : N1098-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14-G-WATER

Quant Time: Jan 12 06:02:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

