

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
 Data File : VN076201.D
 Acq On : 11 Jan 2023 11:40
 Operator : JC\MD
 Sample : PB150155TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150155TB

Quant Time: Jan 12 00:32:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	357966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	562789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	486748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	188772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	183704	47.172	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.340%
35) Dibromofluoromethane	8.177	113	164543	47.211	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	10.570	98	649657	49.855	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.853	95	195854	45.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.760%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	41395	30.835	ug/l	96
37) Ethyl Acetate	7.571	43	41619	11.191	ug/l	99
43) Isopropyl Acetate	8.694	43	209595	34.166	ug/l #	86
95) Naphthalene	15.647	128	11666	3.264	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
Data File : VN076201.D
Acq On : 11 Jan 2023 11:40
Operator : JC\MD
Sample : PB150155TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150155TB

Quant Time: Jan 12 00:32:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

