

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
 Data File : VN076204.D
 Acq On : 11 Jan 2023 12:52
 Operator : JC\MD
 Sample : 01106-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-B

Quant Time: Jan 12 00:34:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	406095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	654273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	581030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	238678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	216282	48.955	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.920%
35) Dibromofluoromethane	8.171	113	193395	47.730	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	10.571	98	761986	50.298	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.600%
62) 4-Bromofluorobenzene	12.853	95	240430	47.917	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	42452	27.875	ug/l	98
20) Methylene Chloride	5.294	84	13355	2.672	ug/l	89
37) Ethyl Acetate	7.571	43	38537	8.914	ug/l	99
43) Isopropyl Acetate	8.694	43	384162	53.866	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
Data File : VN076204.D
Acq On : 11 Jan 2023 12:52
Operator : JC\MD
Sample : 01106-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-B

Quant Time: Jan 12 00:34:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

