

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
 Data File : VN076205.D
 Acq On : 11 Jan 2023 13:16
 Operator : JC\MD
 Sample : 01089-03
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-2-010923

Quant Time: Jan 11 13:38:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	429072	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	670948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	549490	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	230648	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	222874	47.746	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.500%
35) Dibromofluoromethane	8.165	113	193054	46.462	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.920%
50) Toluene-d8	10.565	98	759210	48.870	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.740%
62) 4-Bromofluorobenzene	12.853	95	221636	43.074	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
Data File : VN076205.D
Acq On : 11 Jan 2023 13:16
Operator : JC\MD
Sample : 01089-03
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-2-010923

Quant Time: Jan 11 13:38:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

