

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
 Data File : VN076206.D
 Acq On : 11 Jan 2023 13:40
 Operator : JC\MD
 Sample : 01081-03
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31965

Quant Time: Jan 12 00:34:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	386432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	608664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	504883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	218487	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	193189	45.953	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.900%
35) Dibromofluoromethane	8.171	113	172173	45.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.360%
50) Toluene-d8	10.565	98	685730	48.657	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.320%
62) 4-Bromofluorobenzene	12.853	95	207261	44.402	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.800%
Target Compounds						
95) Naphthalene	15.647	128	32367	4.781	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
Data File : VN076206.D
Acq On : 11 Jan 2023 13:40
Operator : JC\MD
Sample : 01081-03
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31965

Quant Time: Jan 12 00:34:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

