

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
 Data File : VN076211.D
 Acq On : 11 Jan 2023 15:40
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 12 00:37:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	438204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	690239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	590938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	239233	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	227543	47.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.460%
35) Dibromofluoromethane	8.171	113	202337	47.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.680%
50) Toluene-d8	10.571	98	782786	48.979	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.960%
62) 4-Bromofluorobenzene	12.853	95	239688	45.280	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.560%
Target Compounds						
					Qvalue	
59) 2-Hexanone	11.200	43	1655	0.499	ug/l	83
87) 1,3-Dichlorobenzene	13.729	146	1960	0.243	ug/l	89
88) 1,4-Dichlorobenzene	13.729	146	1960	0.237	ug/l	88
91) 1,2-Dichlorobenzene	14.106	146	2088	0.262	ug/l	90
93) 1,2,4-Trichlorobenzene	15.394	180	2583	0.655	ug/l	85
95) Naphthalene	15.635	128	13416	3.163	ug/l	98
96) 1,2,3-Trichlorobenzene	15.841	180	2585	0.702	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
Data File : VN076211.D
Acq On : 11 Jan 2023 15:40
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 12 00:37:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

