

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
 Data File : VN076215.D
 Acq On : 11 Jan 2023 17:59
 Operator : JC\MD
 Sample : 01125-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2023-1-11

Quant Time: Jan 12 00:39:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	400526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	627637	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	546493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	227321	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	205045	47.057	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.120%
35) Dibromofluoromethane	8.171	113	184793	47.543	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.080%
50) Toluene-d8	10.571	98	714491	49.165	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.853	95	225015	46.748	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011123\
Data File : VN076215.D
Acq On : 11 Jan 2023 17:59
Operator : JC\MD
Sample : 01125-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2023-1-11

Quant Time: Jan 12 00:39:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

