

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011124\
 Data File : VN080677.D
 Acq On : 11 Jan 2024 15:40
 Operator : JC\MD
 Sample : P1103-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240110057-03-TRIP-BLANK

Quant Time: Jan 12 05:24:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jan 12 05:20:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	15107	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	128555	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.865	117	139479	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	67278	27.578	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	91.933%
60) 4-Bromofluorobenzene	12.853	95	83211	26.969	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.900%
63) Toluene-d8	10.565	98	184788	30.187	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.633%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011124\
Data File : VN080677.D
Acq On : 11 Jan 2024 15:40
Operator : JC\MD
Sample : P1103-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240110057-03-TRIP-BLANK

Quant Time: Jan 12 05:24:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011124W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jan 12 05:20:11 2024
Response via : Initial Calibration

