

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
 Data File : VN070588.D
 Acq On : 12 Jan 2022 16:01
 Operator : JC/MD
 Sample : N1097-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1D

Quant Time: Jan 13 04:27:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	847038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1459080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1303558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	463964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	590259	48.466	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.940%		
35) Dibromofluoromethane	8.021	113	417043	47.240	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.480%		
50) Toluene-d8	10.441	98	1695903	47.025	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.040%		
62) 4-Bromofluorobenzene	12.731	95	628459	48.248	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.500%		
Target Compounds						
					Qvalue	
16) Acetone	4.288	43	95688	12.103	ug/l	99
18) Methyl Acetate	4.865	43	18228	0.797	ug/l #	71
20) Methylene Chloride	5.101	84	31013	2.835	ug/l #	85
25) 2-Butanone	7.338	43	27748	2.787	ug/l #	88
43) Isopropyl Acetate	8.560	43	131027	4.390	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
Data File : VN070588.D
Acq On : 12 Jan 2022 16:01
Operator : JC/MD
Sample : N1097-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1D

Quant Time: Jan 13 04:27:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

