

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
 Data File : VN070589.D
 Acq On : 12 Jan 2022 16:26
 Operator : JC/MD
 Sample : N1097-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2S

Quant Time: Jan 13 04:27:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	823328	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1437782	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1273245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	439878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	601775	50.835	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.660%	
35) Dibromofluoromethane	8.021	113	424569	48.805	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.600%	
50) Toluene-d8	10.440	98	1732560	48.753	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.500%	
62) 4-Bromofluorobenzene	12.731	95	627456	48.885	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	60381	7.857	ug/l	99
18) Methyl Acetate	4.865	43	26963	1.213	ug/l #	89
20) Methylene Chloride	5.093	84	31127	2.928	ug/l	93
25) 2-Butanone	7.340	43	12391	1.281	ug/l #	89
43) Isopropyl Acetate	8.563	43	142277	4.838	ug/l #	83
68) m/p-Xylenes	11.953	106	8834	0.490	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
Data File : VN070589.D
Acq On : 12 Jan 2022 16:26
Operator : JC/MD
Sample : N1097-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2S

Quant Time: Jan 13 04:27:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

