

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
 Data File : VN070592.D
 Acq On : 12 Jan 2022 17:42
 Operator : JC/MD
 Sample : N1097-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3D

Quant Time: Jan 13 04:27:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	784760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1361648	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1214802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	414050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	605720	53.683	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.360%	
35) Dibromofluoromethane	8.021	113	427295	51.865	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.720%	
50) Toluene-d8	10.440	98	1769510	52.577	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.160%	
62) 4-Bromofluorobenzene	12.731	95	630894	51.901	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 103.800%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	61468	8.392	ug/l	96
18) Methyl Acetate	4.862	43	35059	1.655	ug/l #	70
20) Methylene Chloride	5.095	84	30258	2.986	ug/l #	83
25) 2-Butanone	7.348	43	13772	1.493	ug/l #	89
43) Isopropyl Acetate	8.563	43	153266	5.503	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
Data File : VN070592.D
Acq On : 12 Jan 2022 17:42
Operator : JC/MD
Sample : N1097-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3D

Quant Time: Jan 13 04:27:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

