

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
 Data File : VN070595.D
 Acq On : 12 Jan 2022 18:58
 Operator : JC/MD
 Sample : PB141975ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141975ZHE#02

Quant Time: Jan 13 04:28:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.085	168	815153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1452535	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1283736	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	443934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	637781	54.417	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.840%	
35) Dibromofluoromethane	8.021	113	455464	51.825	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.640%	
50) Toluene-d8	10.440	98	1843731	51.354	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	12.731	95	674812	52.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.080%	
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	57758	7.591	ug/l	94
18) Methyl Acetate	4.859	43	48095	2.185	ug/l	92
20) Methylene Chloride	5.098	84	32594	3.096	ug/l	93
25) 2-Butanone	7.348	43	12608	1.316	ug/l	93
43) Isopropyl Acetate	8.563	43	181372	6.104	ug/l #	80
80) 1,3,5-Trimethylbenzene	13.058	105	17638	0.534	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
Data File : VN070595.D
Acq On : 12 Jan 2022 18:58
Operator : JC/MD
Sample : PB141975ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141975ZHE#02

Quant Time: Jan 13 04:28:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

