

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
 Data File : VN070596.D
 Acq On : 12 Jan 2022 19:23
 Operator : JC/MD
 Sample : PB141975ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141975ZHE#03

Quant Time: Jan 13 04:28:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	736993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1316705	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1168394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	399066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	647739	61.127	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 122.260%			
35) Dibromofluoromethane	8.021	113	458348	57.533	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 115.060%			
50) Toluene-d8	10.440	98	1858388	57.102	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 114.200%			
62) 4-Bromofluorobenzene	12.731	95	673987	57.339	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 114.680%			
Target Compounds						
16) Acetone	4.293	43	94477	13.734	ug/l	94
18) Methyl Acetate	4.862	43	50093	2.517	ug/l #	71
20) Methylene Chloride	5.098	84	57606	6.053	ug/l	87
25) 2-Butanone	7.348	43	17736	2.048	ug/l #	87
43) Isopropyl Acetate	8.566	43	192485	7.147	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
Data File : VN070596.D
Acq On : 12 Jan 2022 19:23
Operator : JC/MD
Sample : PB141975ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141975ZHE#03

Quant Time: Jan 13 04:28:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

