

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
 Data File : VN070597.D
 Acq On : 12 Jan 2022 19:49
 Operator : JC/MD
 Sample : PB141975ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141975ZHE#04

Quant Time: Jan 13 04:28:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	813569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1446193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1278210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	429597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	650945	55.648	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.300%
35) Dibromofluoromethane	8.024	113	464519	53.087	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.180%
50) Toluene-d8	10.440	98	1895872	53.038	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.080%
62) 4-Bromofluorobenzene	12.731	95	672376	52.080	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.160%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	94328	12.422	ug/l	97
18) Methyl Acetate	4.864	43	52059	2.370	ug/l #	88
20) Methylene Chloride	5.098	84	58248	5.544	ug/l	90
25) 2-Butanone	7.343	43	16982	1.776	ug/l	93
43) Isopropyl Acetate	8.563	43	196944	6.658	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011222\
Data File : VN070597.D
Acq On : 12 Jan 2022 19:49
Operator : JC/MD
Sample : PB141975ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141975ZHE#04

Quant Time: Jan 13 04:28:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

