

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
 Data File : VN076224.D
 Acq On : 12 Jan 2023 12:39
 Operator : JC\MD
 Sample : 01126-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-011023

Quant Time: Jan 13 01:57:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	386725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	605049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	532545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	219128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	202109	48.039	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.080%
35) Dibromofluoromethane	8.177	113	180988	48.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.600%
50) Toluene-d8	10.571	98	689549	49.220	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.440%
62) 4-Bromofluorobenzene	12.853	95	217122	46.792	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.580%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	13476	9.292	ug/l #	83
51) 4-Methyl-2-Pentanone	10.453	43	3626	0.909	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
Data File : VN076224.D
Acq On : 12 Jan 2023 12:39
Operator : JC\MD
Sample : 01126-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-011023

Quant Time: Jan 13 01:57:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

