

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
 Data File : VN076225.D
 Acq On : 12 Jan 2023 13:03
 Operator : JC\MD
 Sample : 01130-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230111038-07-TRIP-BLANK

Quant Time: Jan 13 01:58:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	367660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	576456	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	503013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	210315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	191424	47.858	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.720%		
35) Dibromofluoromethane	8.171	113	173448	48.586	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.180%		
50) Toluene-d8	10.571	98	664798	49.807	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.620%		
62) 4-Bromofluorobenzene	12.853	95	207018	46.828	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.660%		
Target Compounds					Qvalue	
16) Acetone	4.454	43	55307	40.112	ug/l	96
30) Chloroform	7.977	83	4438	0.639	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
Data File : VN076225.D
Acq On : 12 Jan 2023 13:03
Operator : JC\MD
Sample : 01130-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230111038-07-TRIP-BLANK

Quant Time: Jan 13 01:58:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

