

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
 Data File : VN076226.D
 Acq On : 12 Jan 2023 13:27
 Operator : JC\MD
 Sample : 01131-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230111038-08-TRIP-BLANK

Quant Time: Jan 13 01:58:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	357581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	570611	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	497323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	204707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	192885	49.583	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.160%
35) Dibromofluoromethane	8.177	113	169576	47.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.980%
50) Toluene-d8	10.571	98	648150	49.057	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.120%
62) 4-Bromofluorobenzene	12.853	95	204167	46.656	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.320%
Target Compounds						
16) Acetone	4.448	43	63437	47.305	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
Data File : VN076226.D
Acq On : 12 Jan 2023 13:27
Operator : JC\MD
Sample : 01131-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230111038-08-TRIP-BLANK

Quant Time: Jan 13 01:58:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

