

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
 Data File : VN076228.D
 Acq On : 12 Jan 2023 14:15
 Operator : JC\MD
 Sample : 01126-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-04-478.6-499.6-011123

Quant Time: Jan 13 01:59:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	366018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	576873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	511631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	213082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	192196	48.267	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.540%		
35) Dibromofluoromethane	8.171	113	172075	48.167	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.340%		
50) Toluene-d8	10.571	98	667502	49.973	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.940%		
62) 4-Bromofluorobenzene	12.853	95	212925	48.129	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.260%		
Target Compounds						
						Qvalue
48) Methyl methacrylate	9.688	41	8517	3.007	ug/l	92
52) Toluene	10.635	92	1204227	126.281	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011223\
Data File : VN076228.D
Acq On : 12 Jan 2023 14:15
Operator : JC\MD
Sample : 01126-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-04-478.6-499.6-011123

Quant Time: Jan 13 01:59:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

