

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011224\
 Data File : VN080684.D
 Acq On : 12 Jan 2024 12:47
 Operator : JC\MD
 Sample : VN0112MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 VN0112MBL01

Quant Time: Jan 12 22:10:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	143769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	333938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	331109	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104465	41.377	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.760%
35) Dibromofluoromethane	8.171	113	87285	46.004	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.000%
50) Toluene-d8	10.565	98	346631	47.912	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.820%
62) 4-Bromofluorobenzene	12.847	95	141515	51.008	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.020%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011224\
 Data File : VN080684.D
 Acq On : 12 Jan 2024 12:47
 Operator : JC\MD
 Sample : VN0112MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0112MBL01

Quant Time: Jan 12 22:10:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

