

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011224\
 Data File : VN080692.D
 Acq On : 12 Jan 2024 16:08
 Operator : JC\MD
 Sample : PB158389ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158389ZHE#05

Quant Time: Jan 12 22:14:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	356530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	112240	44.340	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.680%
35) Dibromofluoromethane	8.165	113	89713	47.543	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.080%
50) Toluene-d8	10.565	98	380438	52.874	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.740%
62) 4-Bromofluorobenzene	12.853	95	159696	57.878	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.760%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	50466	60.725	ug/l	90
20) Methylene Chloride	5.283	84	7728	3.185	ug/l #	92
25) 2-Butanone	7.483	43	8050	5.359	ug/l #	86
43) Isopropyl Acetate	8.688	43	172834	25.086	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011224\
Data File : VN080692.D
Acq On : 12 Jan 2024 16:08
Operator : JC\MD
Sample : PB158389ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158389ZHE#05

Quant Time: Jan 12 22:14:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

