

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011224\
 Data File : VN080693.D
 Acq On : 12 Jan 2024 16:31
 Operator : JC\MD
 Sample : PB158389ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158389ZHE#06

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/15/2024
 Supervised By : Mahesh Dadoda 01/15/2024

Quant Time: Jan 12 22:15:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 06:34:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	146451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	338460	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	351437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	113771	44.237	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.480%		
35) Dibromofluoromethane	8.165	113	91635	47.651	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.300%		
50) Toluene-d8	10.565	98	381988	52.094	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.180%		
62) 4-Bromofluorobenzene	12.847	95	158644	56.418	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	112.840%		
Target Compounds					Qvalue	
16) Acetone	4.424	43	50756	60.113	ug/l	94
20) Methylene Chloride	5.277	84	7578	3.074	ug/l #	94
37) Ethyl Acetate	7.571	43	5530m	1.448	ug/l	
43) Isopropyl Acetate	8.689	43	175577	25.006	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011224\
Data File : VN080693.D
Acq On : 12 Jan 2024 16:31
Operator : JC\MD
Sample : PB158389ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158389ZHE#06

Quant Time: Jan 12 22:15:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 06:34:36 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/15/2024
Supervised By : Mahesh Dadoda 01/15/2024

