

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011322\
 Data File : VN070610.D
 Acq On : 13 Jan 2022 14:33
 Operator : JC/MD
 Sample : N1123-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 220112003-05-TRIP-BLANK

Quant Time: Jan 14 04:53:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	766608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1314917	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1168822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	399798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	628199	56.993	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 113.980%	
35) Dibromofluoromethane	8.021	113	443593	55.756	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 111.520%	
50) Toluene-d8	10.440	98	1756483	54.044	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.080%	
62) 4-Bromofluorobenzene	12.731	95	629298	53.609	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.220%	
Target Compounds						Qvalue
30) Chloroform	7.820	83	16181	0.938	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011322\
Data File : VN070610.D
Acq On : 13 Jan 2022 14:33
Operator : JC/MD
Sample : N1123-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
220112003-05-TRIP-BLANK

Quant Time: Jan 14 04:53:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

