

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
 Data File : VN076237.D
 Acq On : 13 Jan 2023 10:28
 Operator : JC\MD
 Sample : VN0113MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0113MBL01

Quant Time: Jan 16 01:02:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	330808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	518216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	444896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	170101	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	179561	49.894	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.780%
35) Dibromofluoromethane	8.177	113	158647	49.434	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.860%
50) Toluene-d8	10.571	98	601650	50.142	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.280%
62) 4-Bromofluorobenzene	12.853	95	179469	45.159	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011323\
Data File : VN076237.D
Acq On : 13 Jan 2023 10:28
Operator : JC\MD
Sample : VN0113MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0113MBL01

Quant Time: Jan 16 01:02:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

